



TURKISH ACCREDITATION AGENCY

COPY OF THE ACCREDITATION CERTIFICATE

As a Testing Laboratory,

BAYLAN

Ölçü Aletleri Sanayi Ve Ticaret Ltd. Şti.

Atatürk Organize Sanayi Bölgesi 10046 Sok. No:14 Çiğli
35620 İZMİR / TURKEY

The testing laboratories given in Annex are accredited in accordance with TS EN ISO/IEC 17025:2012 for the scopes given, following the assessment conducted by TURKAK.

Accreditation No : AB-0177-T

Accreditation Date : 5 March 2008

Revision Date / Number : 10 December 2014 / 05

This certificate shall remain in force until **11 March 2016**, subject to continuing compliance of the testing laboratories with the standard **TS EN ISO/IEC 17025:2012**, related regulations and requirements.



Dr. H. İbrahim ÇETİN
Deputy Secretary General

Annex of the certificate (Page 1/2)

Accreditation Scope

 Test TS EN ISO/IEC 17025 AB-0177-T	BAYLAN Ölçü Aletleri Sanayi Ve Ticaret Ltd. Şti. Accreditation Nr: AB-0177-T Revision Nr: 05 Date: 10 December 2014	
	As a Testing Laboratory	
	Address: Atatürk Organize Sanayi Bölgesi 10046 Sok. No:14 Çiğli 35620 İZMİR / TÜRKİYE	Phone : 0 232 376 77 57 Fax : 0 232 376 77 59 E-Mail : baylan.lab@gmail.com Website : www.baylansayaclari.com

Tested Materials / Products	Name of Test	Testing Method (National, International standards, in house methods)
Cold And Hot Water Meters (DN 15-200)	Pressure Test	TS 10021 Article 3, ISO 4064-3 Article 6, OIML 49-2:E06 Article 6.2, Directive for Cold Water Meters 75/33 EEC Article 3.3, EN 14154-3 Article 5.2, 2004/22/AT MI001
	Test for determination of error curves according to flow rates	TS 10021 Article 2, ISO 4064-3 Article 5, OIML 49-2:E06 Article 6.3, Directive for Cold Water Meters 75/33 EEC Article 2, EN 14154-3 Article 5.3, 2004/22/AT MI001
	Pressure Loss Test	TS 10021 Article 4, ISO 4064-3 Article 7, OIML 49-2:E04 Article 6.7, Directive for Cold Water Meters 75/33 EEC Article 3.4, EN 14154-3 Article 5.11, 2004/22/AT MI001
	Endurance Test	TS 10021 Article 5, ISO 4064-3 Article 8.1, OIML 49-2:E04 Article 6.9, Directive for Cold Water Meters 75/33 EEC Article 5, EN 14154-3 Article 5.13, 2004/22/AT MI001




Annex of the certificate (Page 2/2)

Accreditation Scope

 <p>TÜRKAK Test TS EN ISO/IEC 17025 AB-0177-T</p>	<p>BAYLAN Ölçü Aletleri Sanayi Ve Ticaret Ltd. Şti.</p> <p>Accreditation Nr: AB-0177-T Revision Nr: 05 Date: 10 December 2014</p>
--	--

Tested Materials / Products	Name of Test	Testing Method (National, International standards, in house methods)
(Cold And Hot Water Meters (DN 15-200), continued)	Reverse Flow Test	EN 14154-3 Article 5.12, ISO 4064-3 Article 5.11, OIML 49-2:E04 Article 6.6, 2004/22/AT MI001
	Zero Flow Test	EN 14154-3 Article 5.5, 2004/22/AT MI001

End of Scope


Dr. H. İbrahim ÇETİN
Acting Secretary General

